TEST CERTIFICATE

Certificate No. RO-50179

Customer: German Tanker Shipping GmbH & Co KG
IMO/ Vessel: 9380477 - Seasprat

Model: Tanksystem UTI / UTImeter
Unit Serial Number: G21282
Display Serial Number: 17799
Tape Serial Number: A9346
Sensor Serial Number: RFC 170

Gärtnerstrasse 81G
D- 25469 HALSTENBEK
Dept.: Technical Services
Tel.:+49 40 - 57148252
Fax:+49 40 - 57148271

Temperature test results:

Test Temperature T_{ref} (°C)	Reading on UTI LCD T_{UTI} (°C)	Deviation T_{ref} - T_{UTI} ≤0.1 (°C)
15,17	15,18	0,01
60,45	60,51	0,06

Calibration Equipment used:
Reference Thermometer Jofra DTI050, SNR: 1658090 with Sensing Probe Jofra STS050, SNR: 595033-01, calibrated 25.10.2019, certificate no. 004693, D-K-15097-01-00, by TES Kalibrierdienst - Pinneberg Lauda Scientific thermoregulated bath, model RE620S, LCK1914

This is to certify that the unit described above has succeeded the following tests and meets the manufacture product specifications.

- Functioning tests
- Grounding
- Detection liquid level (+-2mm)
- Temperature accuracy (+-0.1 °C)
- Gastightness (under 0.3bar)

The above tests have been carried out in complience with requirements and applicable procedures of Honeywell Enraf Tanksystem.

The above results relate only to the items tested. The replacement of Sensor or Display Unit invalidates the current certificate.

Honeywell Enraf Tanksystem SA recommends checking the temperature accuracy once a year and the grounding every 6 months.

Done by: Mike Bonde
Title: Service Technician
01.04.2020

Dept. Technical Services is an official Service Station of Honeywell Enraf Tanksystem.

maritimes cluster norddeutschland
SERVICE POINT KIEL CANAL
Schleswig-Holstein Der echte Norden